



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER OF PATENTS AND TRADEMARKS
Washington, D.C. 20231
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/418,536	10/14/1999	DANIEL J. POWERS	10981567-1	9017

29139 7590 11/26/2002

PHILIPS ELECTRONICS NORTH AMERICAN
580 WHITE PLAINS ROAD
TARRYTOWN, NY 10591

EXAMINER

OROPEZA, FRANCES P

ART UNIT PAPER NUMBER

3762

DATE MAILED: 11/26/2002

Please find below and/or attached an Office communication concerning this application or proceeding.

Advisory Action

Application No.

09/418,536

Applicant(s)

POWERS ET AL.

Examiner

Frances P. Oropeza

Art Unit

3762

--The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

THE REPLY FILED FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE. Therefore, further action by the applicant is required to avoid abandonment of this application. A proper reply to a final rejection under 37 CFR 1.113 may only be either: (1) a timely filed amendment which places the application in condition for allowance; (2) a timely filed Notice of Appeal (with appeal fee); or (3) a timely filed Request for Continued Examination (RCE) in compliance with 37 CFR 1.114.

PERIOD FOR REPLY [check either a) or b)]

- a) ☐ The period for reply expires _____ months from the mailing date of the final rejection.
- b) ☒ The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.
- ONLY CHECK THIS BOX WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).

Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

1. ☐ A Notice of Appeal was filed on _____. Appellant's Brief must be filed within the period set forth in 37 CFR 1.192(a), or any extension thereof (37 CFR 1.191(d)), to avoid dismissal of the appeal.
2. ☒ The proposed amendment(s) will not be entered because:
- (a) ☐ they raise new issues that would require further consideration and/or search (see NOTE below);
- (b) ☐ they raise the issue of new matter (see Note below);
- (c) ☒ they are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or
- (d) ☐ they present additional claims without canceling a corresponding number of finally rejected claims.

NOTE: See Continuation Sheet.

3. ☐ Applicant's reply has overcome the following rejection(s): _____.
4. ☐ Newly proposed or amended claim(s) _____ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).
5. ☐ The a) ☐ affidavit, b) ☐ exhibit, or c) ☐ request for reconsideration has been considered but does NOT place the application in condition for allowance because: _____.
6. ☐ The affidavit or exhibit will NOT be considered because it is not directed SOLELY to issues which were newly raised by the Examiner in the final rejection.
7. ☒ For purposes of Appeal, the proposed amendment(s) a) ☒ will not be entered or b) ☐ will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.

The status of the claim(s) is (or will be) as follows:

Claim(s) allowed: _____.

Claim(s) objected to: _____.

Claim(s) rejected: 1-28.

Claim(s) withdrawn from consideration: _____.

8. ☐ The proposed drawing correction filed on _____ is a) ☐ approved or b) ☐ disapproved by the Examiner.
9. ☐ Note the attached Information Disclosure Statement(s) (PTO-1449) Paper No(s).
10. ☐ Other: _____

[Signature]
PRIMARY EXAMINER
3/11/02
Frances P. Oropeza
11/21/02

Continuation of 2. NOTE:

The amended claims do not place the application in condition for allowance because the prior art of record is deemed to read on the amended independent claims and hence the dependent claims.

Skelton et al. (US 6292692) disclose a medical treatment device and in figure 2 provide a schematic diagram of the device including the control modules "34", "36" and "38". When the system is booted up in the first operation, the system is booted with immediate entry into the basic SAED module "34" that includes ECG monitoring, ECG analysis and a defibrillation module, read as "a defibrillator system to monitor the patient" (c 5, ll 35-39). Modules (36) and (38) include the features of module "34" and make a treatment summary and recorded data available to the user, such that the accumulated recorded ECG data can be displayed on the defibrillator screen (c 6, ll 18-23; c 8, ll 24-26; c11, l 60 - c 12, l 28; figure 4, reference number 98a). When the user enters the manual mode of the device (42), the features of the modules "34", "36" and "38" are supplemented with additional control over the device features and capability (c 6, ll 38-44). Since the "defibrillator system to monitor the patient" is the basic control system of the device applied continuously by all operating modules, the additional module of providing the treatment summary and recorded data on the defibrillator screen is read to be activated while the defibrillator system monitors the patient.

The rejections of record stand.